



Certificate of Classification

Cole-Parmer Product Number: **06409-XX** (xx- represents the size series)

Created/Revised Date **5/14/2012**

Description TUBING MFLEX TYGON LAB

Compound: R-3603

Specification: Tygon® Laboratory

Cole-Parmer certifies that, according to the manufacturer, the product described meets the specifications set forth in this certificate, subject to the disclaimers set forth herein.

All ingredients used in this product are approved under the Code of Federal Regulations 21, subchapter B, Part 175.300-Rubber Articles intended for use in contact with food and drug additives.

This product has been manufactured in accordance with Good Manufacturing Practices.

A lot control system is used to assure traceability to all raw materials.

Cole-Parmer certifies that, according to the most recent information provided by the manufacturer, the product described meets the specifications set forth in this certificate, subject to the disclaimers set forth herein.

This certification applies only to the condition of the above-described product in the unopened package. Cole-Parmer cannot and does not assume any responsibility for any failure of the product to meet this Certificate of Compliance that results from any handling or use subsequent to opening the package.

Cole-Parmer expressly warrants that the product listed above conforms only to the specifications herein, subject to the disclaimers listed above, if any. Except as otherwise set forth in writing by Cole-Parmer, there are no other warranties, express or implied, including without limitation warranties of marketability or fitness for a particular purpose, use or application, that apply to this certificate or the subject product. It is the responsibility of the user to determine the suitability of the subject product for any particular use or application.